

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 10/08/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13200

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:49
INTERNAL STANDARD	VERIFIED	21:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	21:50
BLANK TEST	0.00	21:51
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	21:51
BLANK TEST	0.00	21:52
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	21:52
BLANK TEST	0.00	21:53



565 RJA of Bartlett WESO #162 10/8/13