

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 10/14/2013

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 13200

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:45
INTERNAL STANDARD	VERIFIED	00:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:45
BLANK TEST	0.00	00:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:46
BLANK TEST	0.00	00:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:47
BLANK TEST	0.00	00:48



565 Rph / Barrett WESO #162

10/14/13