SUPERVISOR TEST

```
--------------------- ----
WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208
-------------------------
Date: 10/22/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13200

Nominal = 0.10 g/210L

BLANK TEST                        0.00    00:44
INTERNAL STANDARD             VERIFIED    00:44
REFERENCE STANDARD TEMP         34.00c
REFERENCE STANDARD                0.09    00:44
BLANK TEST                        0.00    00:45
REFERENCE STANDARD TEMP         33.98c
REFERENCE STANDARD                0.09    00:45
BLANK TEST                        0.00    00:46
REFERENCE STANDARD TEMP         33.96c
REFERENCE STANDARD                0.09    00:46
BLANK TEST                        0.00    00:47
```







Sgt WCSO #166 10/22/13