

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 11/04/2013

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 13200

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:05
INTERNAL STANDARD	VERIFIED	18:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:06
BLANK TEST	0.00	18:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	18:07
BLANK TEST	0.00	18:08
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	18:08
BLANK TEST	0.00	18:09



SG RALPH Bartlett WCSO #162 11/04/13