

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 105208

Date: 12/09/2013

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 13260

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:56
INTERNAL STANDARD	VERIFIED	01:56
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:57
BLANK TEST	0.00	01:57
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:58
BLANK TEST	0.00	01:58
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:59
BLANK TEST	0.00	01:59



SOT Dept / J Bartlett W850 # 166 12-09-13