

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 12/16/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13260

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:49
INTERNAL STANDARD	VERIFIED	20:49
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	20:49
BLANK TEST	0.00	20:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	20:50
BLANK TEST	0.00	20:51
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	20:51
BLANK TEST	0.00	20:52



SGR RJB / Bartlett WCSO #162 12-16-13