

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 12/31/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13260

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:07
INTERNAL STANDARD	VERIFIED	23:07
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	23:08
BLANK TEST	0.00	23:08
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	23:09
BLANK TEST	0.00	23:09
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	23:10
BLANK TEST	0.00	23:11



SGS *RJ Bartlett*

WCSO #102

12-31-13