

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 01/02/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12190

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:26
INTERNAL STANDARD	VERIFIED	03:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:26
BLANK TEST	0.00	03:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:27
BLANK TEST	0.00	03:28
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.09	03:28
BLANK TEST	0.00	03:29



SGS *Ralph J Bartlett* Wcso #162 01-02-13