

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 01/15/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12190

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:54
INTERNAL STANDARD	VERIFIED	00:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:55
BLANK TEST	0.00	00:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:56
BLANK TEST	0.00	00:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:57
BLANK TEST	0.00	00:58



SGS Rusty Bartlett w/cio #162 01-15-13