

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 01/21/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12190

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:08
INTERNAL STANDARD	VERIFIED	03:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:08
BLANK TEST	0.00	03:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:09
BLANK TEST	0.00	03:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:10
BLANK TEST	0.00	03:11



SGT Ralph Bartlett

WCSO #162

01-21-13