

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 01/30/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12230

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:35
INTERNAL STANDARD	VERIFIED	01:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:36
BLANK TEST	0.00	01:37
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:37
BLANK TEST	0.00	01:38
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:38
BLANK TEST	0.00	01:39



SGS *Ralph J Bartlett* WCSO #162 01-30-13