

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 02/12/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12230

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:40
INTERNAL STANDARD	VERIFIED	00:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:41
BLANK TEST	0.00	00:41
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:42
BLANK TEST	0.00	00:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:43
BLANK TEST	0.00	00:44



SGS By *Ralph Bartlett* WCSO SHERIFF #162 02-12-13