

SUPERVISOR TEST




WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 02/26/2013

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 12230

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:33
INTERNAL STANDARD	VERIFIED	00:33
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:33
BLANK TEST	0.00	00:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:34
BLANK TEST	0.00	00:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:35
BLANK TEST	0.00	00:36



SGS R.J. Bartlett

WCSO

SHERIFF #162

02-26-13