

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 03/04/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12230

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:02
INTERNAL STANDARD	VERIFIED	02:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:02
BLANK TEST	0.00	02:03
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	02:03
BLANK TEST	0.00	02:04
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:05
BLANK TEST	0.00	02:05



SGS Rpt 1 Bureau

WCSO # 162

07-04-13