

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 03/12/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12230

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:44
INTERNAL STANDARD	VERIFIED	00:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:45
BLANK TEST	0.00	00:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:46
BLANK TEST	0.00	00:47
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.10	00:47
BLANK TEST	0.00	00:48



SG- Rpt 7 Bartlett WCSD SHERIFF #166 03-12-13