

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 03/17/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12230

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:37
INTERNAL STANDARD	VERIFIED	11:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:38
BLANK TEST	0.00	11:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:39
BLANK TEST	0.00	11:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	11:40
BLANK TEST	0.00	11:41



SGT Ralph Bartlett

WCSO #162

03-17-13