

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 04/09/2013

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 13050

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:10
INTERNAL STANDARD	VERIFIED	01:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:10
BLANK TEST	0.00	01:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:11
BLANK TEST	0.00	01:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:12
BLANK TEST	0.00	01:13



565 Dept of Bartlett WCSO #162 04-09-13