

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 04/15/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13050

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:02
INTERNAL STANDARD	VERIFIED	00:02
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:03
BLANK TEST	0.00	00:03
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:04
BLANK TEST	0.00	00:05
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	00:05
BLANK TEST	0.00	00:06



SGT *Ralph Bartlett* WCR #162 04-15-13