

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 04/23/2013

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 13050

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:03
INTERNAL STANDARD	VERIFIED	02:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:04
BLANK TEST	0.00	02:04
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:05
BLANK TEST	0.00	02:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:06
BLANK TEST	0.00	02:06



SGT *RJ Bartlett* WESO #102 04-23-13