

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 05/08/2013

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 13050

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:45
INTERNAL STANDARD	VERIFIED	06:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	06:46
BLANK TEST	0.00	06:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:47
BLANK TEST	0.00	06:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:48
BLANK TEST	0.00	06:49



*SOS Ralph Bartlett*

*WCSO # 162*

*05-08-13*