

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 06/03/2013

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 13090

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:44
INTERNAL STANDARD	VERIFIED	21:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	21:45
BLANK TEST	0.00	21:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	21:46
BLANK TEST	0.00	21:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:47
BLANK TEST	0.00	21:47



565 *Ralph Bartlett* WCO #162 06-03-13