

SUPERVISOR TEST




WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 06/18/2013

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 13090

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:24
INTERNAL STANDARD	VERIFIED	00:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:24
BLANK TEST	0.00	00:25
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:25
BLANK TEST	0.00	00:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:26
BLANK TEST	0.00	00:27



SGS Dept of Justice WOOD #162 06-18-13