

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 06/24/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13090

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:47
INTERNAL STANDARD	VERIFIED	02:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:48
BLANK TEST	0.00	02:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:49
BLANK TEST	0.00	02:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:50
BLANK TEST	0.00	02:51



505 *Ralph Bartlett* WESO #162 06-24-13