

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 07/02/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13090

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:46
INTERNAL STANDARD	VERIFIED	01:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:47
BLANK TEST	0.00	01:48
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	01:48
BLANK TEST	0.00	01:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:49
BLANK TEST	0.00	01:50



SGS Dept of Justice Waco #122 07-02-13