

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 07/17/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13090

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:40
INTERNAL STANDARD	VERIFIED	01:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:40
BLANK TEST	0.00	01:41
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:41
BLANK TEST	0.00	01:42
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:42
BLANK TEST	0.00	01:43



SGT *RJ Bartlett* WISO #162 07-17-13