

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 07/24/2013

Operator Name: JOSHUA M LOPEZ
Permit #: 34728
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13090

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:45
INTERNAL STANDARD	VERIFIED	20:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	20:45
BLANK TEST	0.00	20:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:47
BLANK TEST	0.00	20:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:48
BLANK TEST	0.00	20:48

P.O. Jm Lopez #199

