

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 07/30/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13140

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:32
INTERNAL STANDARD	VERIFIED	05:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:32
BLANK TEST	0.00	05:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	05:33
BLANK TEST	0.00	05:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	05:34
BLANK TEST	0.00	05:35



SUS Dept of Bartlett WESO #102 07-30-13