

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 09/16/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13140

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:30
INTERNAL STANDARD	VERIFIED	02:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:31
BLANK TEST	0.00	02:31
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:32
BLANK TEST	0.00	02:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:33
BLANK TEST	0.00	02:33



SGT *Ralph Bartlett* WCSO #162 09-16-13