

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 09/30/2013

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 13200

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:06
INTERNAL STANDARD	VERIFIED	01:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:07
BLANK TEST	0.00	01:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:08
BLANK TEST	0.00	01:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:09
BLANK TEST	0.00	01:10



SGT Sgt J Bartlett W250#162

09/30/13