

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/14/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13200

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:16
INTERNAL STANDARD	VERIFIED	01:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:16
BLANK TEST	0.00	01:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:17
BLANK TEST	0.00	01:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:18
BLANK TEST	0.00	01:19



SGT R. J. Bartlett WESO #162 10-14-13