

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER cmt: 144306

Date: 10/22/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13200

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:12
INTERNAL STANDARD	VERIFIED	01:12
REFERENCE STANDARD TEMP	33.56c	
REFERENCE STANDARD	0.09	01:13
BLANK TEST	0.00	01:13
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	01:14
BLANK TEST	0.00	01:14
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:15
BLANK TEST	0.00	01:16



505 *RJ Bartlett*

WCSO #120

10/22/13