

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/28/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13200

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:38
INTERNAL STANDARD	VERIFIED	00:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:39
BLANK TEST	0.00	00:39
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:40
BLANK TEST	0.00	00:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:41
BLANK TEST	0.00	00:41



SGT Sgt J Bartlett WESO #142 10-28-13