

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 11/25/2013

Operator Name: CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13200

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:53
INTERNAL STANDARD	VERIFIED	09:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:54
BLANK TEST	0.00	09:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:55
BLANK TEST	0.00	09:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	09:56
BLANK TEST	0.00	09:57

Charles B. Geisler
11/25/13 1000

