

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 12/16/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13260

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:22
INTERNAL STANDARD	VERIFIED	22:22
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	22:23
BLANK TEST	0.00	22:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:24
BLANK TEST	0.00	22:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:25
BLANK TEST	0.00	22:26



SGT R.J. Bartlett WSO #100 12-16-13