

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 12/21/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13260

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:14
INTERNAL STANDARD	VERIFIED	02:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:15
BLANK TEST	0.00	02:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:16
BLANK TEST	0.00	02:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:17
BLANK TEST	0.00	02:18



565 Rpt of Results WOSO #162 12-21-13