

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 12/27/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13260

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:26
INTERNAL STANDARD	VERIFIED	01:26
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:27
BLANK TEST	0.00	01:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:28
BLANK TEST	0.00	01:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:29
BLANK TEST	0.00	01:30



SGS *RJ Bartlett*

WCISO #102 12-27-13