

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 12/31/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13260

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:35
INTERNAL STANDARD	VERIFIED	22:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:35
BLANK TEST	0.00	22:36
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	22:36
BLANK TEST	0.00	22:37
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	22:37
BLANK TEST	0.00	22:38



SGS *RJA / Bartlett* WCSO #162 12-31-13