

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 01/14/2014

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13260

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:03
INTERNAL STANDARD	VERIFIED	01:03
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:03
BLANK TEST	0.00	01:04
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	01:04
BLANK TEST	0.00	01:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:05
BLANK TEST	0.00	01:06

