

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 01/24/2014

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 13260

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:02
INTERNAL STANDARD	VERIFIED	01:02
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:02
BLANK TEST	0.00	01:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:04
BLANK TEST	0.00	01:04
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	01:05
BLANK TEST	0.00	01:05

