

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 01/28/2014

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13260

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:49
INTERNAL STANDARD	VERIFIED	00:49
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	00:50
BLANK TEST	0.00	00:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:51
BLANK TEST	0.00	00:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:52
BLANK TEST	0.00	00:53



SO: R J Bartlett WCO #162 01-28-14