

SUPERVISOR TEST




-----  
 WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 105208  
 -----

Date: 02/17/2014

Operator Name: RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 13320

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:07
INTERNAL STANDARD	VERIFIED	00:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:08
BLANK TEST	0.00	00:08
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	00:09
BLANK TEST	0.00	00:10
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	00:10
BLANK TEST	0.00	00:11



SG: RJA Bartlett weso #162 02-17-14