

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 02/25/2014

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 13320

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:27
INTERNAL STANDARD	VERIFIED	02:27
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:27
BLANK TEST	0.00	02:28
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:28
BLANK TEST	0.00	02:29
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	02:29
BLANK TEST	0.00	02:30



505 BJA, Bartlett WCSO #162 02-25-14