

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 03/11/2014

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13320

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:56
INTERNAL STANDARD	VERIFIED	02:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:57
BLANK TEST	0.00	02:57
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	02:58
BLANK TEST	0.00	02:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:59
BLANK TEST	0.00	03:00



SGT RALPH BARTLETT WESO #162 03-11-14