

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 04/01/2014

Operator Name: JOSHUA M LOPEZ
Permit #: 34728
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14040

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:31
INTERNAL STANDARD	VERIFIED	09:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:32
BLANK TEST	0.00	09:32
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	09:33
BLANK TEST	0.00	09:33
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	09:34
BLANK TEST	0.00	09:34

J.C. Lopez 299

