

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 04/22/2014

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:50
INTERNAL STANDARD	VERIFIED	22:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:51
BLANK TEST	0.00	22:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:52
BLANK TEST	0.00	22:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:53
BLANK TEST	0.00	22:54



SGS RJA/Bartlett WSP #162 04-22-14