

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 05/15/2014

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14040

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:11
INTERNAL STANDARD	VERIFIED	01:11
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	01:12
BLANK TEST	0.00	01:12
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:13
BLANK TEST	0.00	01:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:14
BLANK TEST	0.00	01:14



565 Bly Juntas weso #162 05-15-14