

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 05/19/2014

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 14040

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:29
INTERNAL STANDARD	VERIFIED	21:29
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	21:29
BLANK TEST	0.00	21:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:30
BLANK TEST	0.00	21:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:31
BLANK TEST	0.00	21:32



SGS Ralph J Bartlett WSC #162 05-19-14