




SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 06/17/2014

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14080
Nominal = 0.10 g/210L

BLANK TEST	0.00	00:54
INTERNAL STANDARD	VERIFIED	00:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:54
BLANK TEST	0.00	00:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:55
BLANK TEST	0.00	00:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:56
BLANK TEST	0.00	00:57



See Page 1, Results WCSO #162 06-17-14