

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 07/01/2014

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14080

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:42
INTERNAL STANDARD	VERIFIED	01:42
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:42
BLANK TEST	0.00	01:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:43
BLANK TEST	0.00	01:44
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	01:44
BLANK TEST	0.00	01:45



565 Ralph Bartlett WGSO #12 07-01-14