

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/04/2014

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14140

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:01
INTERNAL STANDARD	VERIFIED	00:01
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:01
BLANK TEST	0.00	00:02
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	00:02
BLANK TEST	0.00	00:03
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:03
BLANK TEST	0.00	00:04



SGS RJA Bartlett WCO #162 08-04-14